
C35 Plus Hyperextension Brace

To open the lock for the lower bar on the hyperextension brace. Unlock the screw under the cover (with the allen wrench that is supplied with the brace). This will make the lower bar move with the patient and flex during movement.







SVERIGE Tel 042-25 27 00 E-post info@camp.se DANMARK 43 96 66 99 info@camp.dk FINLAND 09-350 76 30 info@camp.fi NORGE 23 23 31 20 info@camp.no